
Midmark M11 - v1.0.5
Total Cycles: 4439

Name: RITTER2
Sterilizer ID: V2313228
03/20/2025 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.4 F 0.0
2:00 109.6 F 0.9
4:00 120.4 F 3.1
6:00 143.6 F 6.4
8:00 178.5 F 10.1
10:00 195.4 F 8.9
12:00 213.1 F 7.6
14:00 229.5 F 11.7
16:00 229.6 F 7.8
18:00 239.3 F 13.0
20:00 239.8 F 11.1
22:00 251.8 F 17.1
24:00 262.6 F 23.4
26:00 272.6 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 272.8 F 30.4
1:00 272.0 F 30.4
2:00 272.1 F 30.4
3:00 272.4 F 30.5
4:00 272.3 F 30.7

Min 271.7 F 30.3
Max 272.8 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 188.4 F 0.0
10:00 199.3 F 0.0
15:00 195.3 F 0.0
20:00 192.0 F 0.0
25:00 191.4 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:09
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:47
CYCLE COMPLETE

03/20/2025 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____