
Midmark M11 - v1.0.5
Total Cycles: 1984

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.2 F 0.0
2:00 167.5 F 0.4
4:00 171.5 F 3.9
6:00 194.3 F 9.1
8:00 219.6 F 7.7
10:00 232.1 F 10.4
12:00 231.8 F 7.7
14:00 242.6 F 11.9
16:00 245.8 F 13.4
18:00 258.7 F 20.4
20:00 270.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.4 F 29.5
2:00 271.6 F 28.8
3:00 271.5 F 29.2
4:00 272.2 F 29.4

Min 271.0 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 214.6 F 0.0
10:00 226.0 F 0.0
15:00 228.9 F 0.0
20:00 228.6 F 0.0
25:00 229.9 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

01/31/2023 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____