

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3488

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2024 01:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.9 F 0.0  
2:00 161.8 F 1.9  
4:00 176.1 F 5.6  
6:00 201.3 F 10.9  
8:00 219.9 F 5.7  
10:00 231.4 F 11.2  
12:00 232.5 F 8.8  
14:00 239.0 F 9.9  
16:00 247.7 F 14.6  
18:00 260.0 F 21.6  
20:00 271.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.7 F 30.1  
2:00 272.7 F 30.2  
3:00 271.9 F 30.1  
4:00 271.7 F 29.6  
  
Min 270.8 F 29.0  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.4 F 0.0  
5:00 236.5 F 0.0  
10:00 189.0 F 0.0  
15:00 184.0 F 0.0  
20:00 181.4 F 0.0  
25:00 179.3 F 0.0  
30:00 170.5 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:35  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:10  
CYCLE COMPLETE

03/28/2024 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_