

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001035  
12:29 13/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:59  
STERILIZING 5:17  
136.0 C 220kPa 5:17  
136.1 C 221kPa 5:47  
136.1 C 221kPa 6:17  
136.1 C 223kPa 6:47  
136.3 C 225kPa 7:17  
136.4 C 225kPa 7:47  
136.4 C 224kPa 8:17  
136.3 C 223kPa 8:47  
136.3 C 223kPa 8:47  
VENTING 8:47  
AIR DRYING 9:45  
CYCLE COMPLETE 24:45  
-----

Digital Signature #  
432F866B2067D041  
-----