

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1690

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2022 03:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
2:00 185.1 F 1.4  
4:00 194.8 F 6.1  
6:00 212.9 F 11.4  
8:00 224.0 F 6.9  
10:00 237.5 F 12.6  
12:00 238.8 F 10.8  
14:00 238.0 F 9.9  
16:00 252.1 F 16.9  
18:00 264.1 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.4 F 29.4  
2:00 271.8 F 29.4  
3:00 272.3 F 29.4  
4:00 271.5 F 29.1  
  
Min 271.5 F 28.9  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
5:00 227.9 F 0.0  
10:00 235.8 F 0.0  
15:00 229.0 F 0.0  
20:00 226.4 F 0.0  
25:00 225.5 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 19:55  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

11/07/2022 04:14 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_