
Midmark M11 - v1.0.0
Total Cycles: 2650

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 79.9 F 0.0
4:00 86.0 F 0.0
6:00 112.6 F 0.5
8:00 142.5 F 1.9
10:00 172.3 F 4.2
12:00 193.8 F 7.4
14:00 206.7 F 11.1
16:00 217.8 F 4.8
18:00 227.7 F 8.3
20:00 237.4 F 12.5
22:00 236.3 F 9.2
24:00 239.7 F 9.9
26:00 245.8 F 13.1
28:00 257.3 F 19.2
30:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.1
2:00 271.7 F 28.2
3:00 272.2 F 28.9
4:00 272.3 F 28.7

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 227.9 F 0.0
10:00 223.9 F 0.0
15:00 230.5 F 0.0
20:00 229.4 F 0.0
25:00 230.8 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:25
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

12/07/2022 10:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____