
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002317
8:50 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:44
STERILIZING 3:50
136.0 C 220kPa 3:50
136.3 C 224kPa 4:21
136.4 C 224kPa 4:51
136.0 C 221kPa 5:21
136.4 C 224kPa 5:51
136.4 C 224kPa 6:21
136.3 C 223kPa 6:51
136.4 C 223kPa 7:21
136.4 C 223kPa 7:21
VENTING 7:21
AIR DRYING 8:11
CYCLE COMPLETE 25:15

Digital Signature #
2769DBF41C527EC4
