
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001085
17:00 18/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:43
STERILIZING 5:25
136.5 C 224kPa 5:25
136.2 C 224kPa 5:55
136.3 C 224kPa 6:25
136.3 C 224kPa 6:55
136.3 C 225kPa 7:25
136.4 C 225kPa 7:55
136.3 C 224kPa 8:25
136.3 C 224kPa 8:55
136.3 C 224kPa 8:55
VENTING 8:56
AIR DRYING 10:02
CYCLE COMPLETE 25:03

Digital Signature #
B4DC8D604CC5931B
