
Midmark M11 - v1.0.5
Total Cycles: 3731

Name: RITTER2
Sterilizer ID: V2313228
06/24/2024 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 82.6 F 0.2
4:00 80.4 F 1.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 89F
CHAMBER PRESS 1.7PSI
** ITEMS NOT STERILE **

06/24/2024 02:06 PM