

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2850

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2023 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.6 F 0.0  
2:00 172.8 F 3.3  
4:00 194.5 F 7.2  
6:00 215.7 F 12.0  
8:00 219.0 F 6.2  
10:00 228.1 F 11.3  
12:00 232.6 F 7.6  
14:00 244.6 F 13.4  
16:00 242.2 F 11.7  
18:00 255.2 F 18.2  
20:00 266.5 F 25.1  
22:00 272.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.2  
1:00 272.4 F 29.3  
2:00 272.5 F 29.7  
3:00 272.0 F 29.3  
4:00 272.4 F 29.4

Min 271.9 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 213.6 F 0.0  
10:00 221.7 F 0.0  
15:00 236.8 F 0.0  
20:00 235.3 F 0.0  
25:00 228.5 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:22  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:44  
CYCLE COMPLETE

02/02/2023 03:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_