
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007000
7:49 28/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 4:05
STERILIZING 5:04
136.0 C 220kPa 5:04
136.4 C 224kPa 5:34
136.4 C 222kPa 6:04
136.3 C 222kPa 6:34
136.4 C 223kPa 7:04
136.4 C 223kPa 7:34
136.4 C 222kPa 8:04
136.3 C 222kPa 8:34
136.3 C 222kPa 8:34
VENTING 8:34
AIR DRYING 9:27
CYCLE COMPLETE 9:40

Digital Signature #
1534B031FA86EAFE
