
Midmark M11 - v1.0.5
Total Cycles: 509

Name: RITTER2
Sterilizer ID: V2313228
11/01/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 130.0 F 1.4
4:00 146.9 F 4.3
6:00 169.5 F 7.8
8:00 195.6 F 11.9
10:00 210.2 F 4.8
12:00 224.0 F 9.5
14:00 231.0 F 9.8
16:00 239.5 F 11.9
18:00 238.6 F 10.5
20:00 251.4 F 16.7
22:00 262.2 F 23.1
24:00 272.4 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.7 F 29.9
2:00 271.5 F 29.3
3:00 272.6 F 30.0
4:00 272.0 F 29.4

Min 271.5 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 207.2 F 0.0
10:00 212.1 F 0.0
15:00 208.8 F 0.0
20:00 191.1 F 0.0
25:00 175.3 F 0.0
30:00 164.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

11/01/2021 11:02 AM
ACCEPTED: X
REJECTED:
Operator: U01