
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004870
11:09 20/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:16
136.0 C 220kPa 3:16
135.8 C 217kPa 3:46
136.3 C 223kPa 4:16
136.4 C 222kPa 4:46
136.4 C 224kPa 5:16
136.4 C 224kPa 5:46
136.4 C 225kPa 6:16
136.3 C 222kPa 6:46
136.3 C 222kPa 6:46
VENTING 6:46
AIR DRYING 7:35
CYCLE COMPLETE 11:51

Digital Signature #
9CA0D58C8289D247
