
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002624
14:20 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:57
STERILIZING 2:54
136.0 C 220kPa 2:54
136.4 C 225kPa 3:24
136.3 C 223kPa 3:54
136.3 C 225kPa 4:24
136.3 C 225kPa 4:54
136.3 C 225kPa 5:24
136.4 C 226kPa 5:54
136.4 C 226kPa 6:24
136.4 C 226kPa 6:24
VENTING 6:24
AIR DRYING 7:17
CYCLE COMPLETE 29:24

Digital Signature #
19177CC56C3E805D
