
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000864
16:03 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:57
STERILIZING 4:06
136.3 C 220kPa 4:06
136.3 C 233kPa 4:36
136.1 C 231kPa 5:06
136.3 C 232kPa 5:36
136.4 C 230kPa 6:06
136.0 C 230kPa 6:36
136.4 C 233kPa 7:06
136.2 C 230kPa 7:36
136.2 C 230kPa 7:36
VENTING 7:36
AIR DRYING 8:27
CYCLE COMPLETE 21:49

Digital Signature #
CD9A27180A97289C
