
Midmark M11 - v1.0.5
Total Cycles: 2005

Name: RITTER2
Sterilizer ID: V2313228
02/03/2023 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	78.4 F	0.0
2:00	79.5 F	0.0
4:00	86.8 F	0.0
6:00	121.1 F	0.0
8:00	152.6 F	3.7
10:00	188.6 F	7.7
12:00	204.6 F	12.6
14:00	218.5 F	6.3
16:00	230.2 F	10.5
18:00	233.7 F	8.0
20:00	242.3 F	12.3
22:00	241.0 F	10.9
24:00	254.4 F	17.7
26:00	266.3 F	25.0

STERILIZING

mm:ss	Degrees	PSI
0:00	271.7 F	28.6
1:00	271.9 F	28.9
2:00	272.4 F	29.3
3:00	272.8 F	29.3
4:00	272.8 F	29.3

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

C106: DOOR MODE
STOP PRESSED
CHAMBER TEMP 239F
CHAMBER PRESS 0.0PSI

** ITEMS NOT STERILE **

02/03/2023 10:27 AM