

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001893  
12:07 25/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 1:58  
STERILIZING 3:00  
136.4 C 224kPa 3:00  
136.2 C 223kPa 3:31  
136.2 C 224kPa 4:01  
136.4 C 226kPa 4:31  
136.4 C 226kPa 5:01  
136.4 C 226kPa 5:31  
136.4 C 225kPa 6:01  
136.4 C 226kPa 6:31  
136.4 C 226kPa 6:31  
VENTING 6:31  
AIR DRYING 7:21  
CYCLE COMPLETE 7:32

-----  
Digital Signature #  
5E405834779EF126  
-----