
Midmark M11 - v1.0.5
Total Cycles: 2556

Name: RITTER2
Sterilizer ID: V2313228
06/23/2023 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 171.7 F 1.7
4:00 176.7 F 5.4
6:00 201.2 F 10.4
8:00 219.2 F 5.9
10:00 230.1 F 10.1
12:00 233.1 F 8.6
14:00 243.0 F 13.1
16:00 242.2 F 11.9
18:00 254.6 F 18.3
20:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.8
2:00 271.8 F 29.1
3:00 271.6 F 29.5
4:00 272.4 F 29.7

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 201.9 F 0.0
10:00 213.2 F 0.0
15:00 214.2 F 0.0
20:00 214.2 F 0.0
25:00 214.0 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

06/23/2023 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____