

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1235

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2022 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
2:00 173.0 F 2.7  
4:00 175.2 F 6.9  
6:00 200.9 F 11.7  
8:00 215.5 F 6.2  
10:00 230.1 F 11.2  
12:00 231.1 F 7.9  
14:00 243.5 F 13.6  
16:00 242.5 F 12.2  
18:00 255.0 F 18.7  
20:00 266.1 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.1 F 29.5  
2:00 272.3 F 30.0  
3:00 271.4 F 29.3  
4:00 272.3 F 29.8

Min 271.3 F 29.3  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 206.6 F 0.0  
10:00 216.6 F 0.0  
15:00 220.0 F 0.0  
20:00 229.1 F 0.0  
25:00 219.5 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:40  
CYCLE COMPLETE

06/28/2022 01:13 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02