

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006464  
8:16 16/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:26  
PRESSURIZING 1:30  
STERILIZING 2:36  
136.4 C 221kPa 2:36  
136.4 C 223kPa 3:06  
136.4 C 218kPa 3:36  
136.3 C 221kPa 4:06  
136.4 C 221kPa 4:36  
136.4 C 219kPa 5:06  
136.4 C 221kPa 5:36  
136.4 C 221kPa 6:06  
136.4 C 221kPa 6:06  
VENTING 6:06  
AIR DRYING 6:56  
CYCLE COMPLETE 7:35

-----  
Digital Signature #  
60AA88CA2AB59347  
-----