

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1207

Name: RITTER2  
Sterilizer ID: V2313228  
06/21/2022 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.5 F 0.0  
2:00 155.8 F 2.4  
4:00 169.3 F 6.7  
6:00 191.7 F 11.6  
8:00 213.3 F 5.4  
10:00 226.6 F 10.4  
12:00 230.3 F 8.0  
14:00 240.8 F 12.6  
16:00 240.0 F 11.1  
18:00 252.8 F 17.5  
20:00 263.9 F 24.1  
22:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.6 F 30.0  
2:00 271.4 F 29.5  
3:00 272.6 F 30.1  
4:00 271.9 F 29.6

Min 271.4 F 29.3  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 205.0 F 0.0  
10:00 215.0 F 0.0  
15:00 213.7 F 0.0  
20:00 205.5 F 0.0  
25:00 201.4 F 0.0  
30:00 201.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:55  
CYCLE COMPLETE

06/21/2022 12:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_