

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 001089  
13:53 20/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:18  
STERILIZING 4:29  
136.5 C 226kPa 4:29  
136.3 C 225kPa 4:59  
136.3 C 225kPa 5:29  
136.1 C 224kPa 5:59  
136.2 C 225kPa 6:29  
136.4 C 222kPa 6:59  
136.4 C 224kPa 7:29  
136.4 C 225kPa 7:59  
136.4 C 225kPa 7:59  
VENTING 7:59  
AIR DRYING 8:51  
CYCLE COMPLETE 23:51  
-----

Digital Signature #  
319BA1B56ABB6DBC  
-----