
Midmark M11 - v1.0.5
Total Cycles: 69

Name: RITTER2
Sterilizer ID: V2313228
06/07/2021 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 181.4 F 3.1
4:00 202.4 F 8.8
6:00 220.2 F 7.7
8:00 235.3 F 12.1
10:00 237.9 F 11.0
12:00 239.1 F 10.6
14:00 253.8 F 17.9
16:00 265.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.5 F 29.6
2:00 271.9 F 29.6
3:00 272.0 F 29.2
4:00 272.5 F 29.9

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 220.2 F 0.0
10:00 243.7 F 0.0
15:00 236.4 F 0.0
20:00 227.3 F 0.0
25:00 226.1 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 17:25
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:34
CYCLE COMPLETE

06/07/2021 11:07 AM
ACCEPTED: X
REJECTED:
Operator: U01