
Midmark M11 - v1.0.5
Total Cycles: 1027

Name: RITTER2
Sterilizer ID: V2313228
04/25/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.5 F 0.0
2:00 146.7 F 2.1
4:00 157.0 F 5.4
6:00 190.0 F 9.9
8:00 216.2 F 6.6
10:00 227.1 F 9.0
12:00 237.2 F 10.9
14:00 238.2 F 11.4
16:00 237.9 F 10.0
18:00 251.1 F 16.5
20:00 262.4 F 23.1
22:00 272.6 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.6
1:00 272.6 F 30.3
2:00 271.7 F 29.5
3:00 272.7 F 30.2
4:00 271.6 F 29.6

Min 271.5 F 29.5
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 204.9 F 0.0
10:00 217.9 F 0.0
15:00 218.9 F 0.0
20:00 221.9 F 0.0
25:00 231.5 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

04/25/2022 02:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____