

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5255

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2025 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.8 F 0.0  
2:00 117.0 F 0.0  
4:00 139.8 F 0.1  
6:00 170.6 F 2.4  
8:00 194.2 F 5.6  
10:00 204.6 F 9.1  
12:00 217.1 F 13.1  
14:00 223.7 F 6.5  
16:00 233.9 F 10.7  
18:00 232.7 F 7.5  
20:00 243.7 F 12.6  
22:00 241.4 F 10.9  
24:00 253.7 F 17.0  
26:00 264.3 F 23.3  
28:00 271.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.6 F 29.0  
2:00 272.3 F 28.5  
3:00 271.5 F 28.1  
4:00 272.2 F 28.8

Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 199.4 F 0.0  
10:00 201.7 F 0.0  
15:00 190.4 F 0.0  
20:00 192.2 F 0.0  
25:00 199.2 F 0.0  
30:00 197.0 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 28:09  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:06:46  
CYCLE COMPLETE

01/17/2025 03:30 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_