
Midmark M11 - v1.0.5
Total Cycles: 2319

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 186.0 F 1.8
4:00 210.0 F 6.8
6:00 228.4 F 12.6
8:00 233.8 F 10.0
10:00 234.7 F 8.8
12:00 239.5 F 10.4
14:00 249.4 F 15.7
16:00 262.5 F 23.1
18:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.0 F 29.4
2:00 272.6 F 29.9
3:00 271.5 F 29.2
4:00 272.3 F 29.6

Min 271.4 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 226.9 F 0.0
10:00 233.0 F 0.0
15:00 231.9 F 0.0
20:00 227.9 F 0.0
25:00 228.6 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:56:18
CYCLE COMPLETE

05/01/2023 12:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____