
Midmark M11 - v1.0.0
Total Cycles: 4602

Name: RITTER 1
Sterilizer ID: V2178198
06/03/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.6 F 0.0
2:00 117.8 F 1.3
4:00 128.7 F 3.6
6:00 155.9 F 6.7
8:00 180.5 F 10.6
10:00 205.0 F 3.8
12:00 216.5 F 7.0
14:00 230.8 F 11.4
16:00 229.9 F 7.4
18:00 242.7 F 12.9
20:00 240.4 F 10.8
22:00 252.7 F 16.9
24:00 263.6 F 23.1
26:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.7 F 29.3
2:00 271.5 F 28.4
3:00 272.3 F 29.3
4:00 272.2 F 28.9

Min 271.2 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 205.0 F 0.0
10:00 205.2 F 0.0
15:00 218.0 F 0.0
20:00 226.5 F 0.0
25:00 226.5 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:16
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

06/03/2024 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____