
Midmark M11 - v1.0.5
Total Cycles: 2798

Name: RITTER2
Sterilizer ID: V2313228
08/25/2023 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.8 F 0.0
2:00 109.9 F 0.0
4:00 132.0 F 0.0
6:00 195.1 F 0.0
8:00 201.5 F 1.6
10:00 209.0 F 4.6
12:00 223.7 F 7.7
14:00 231.7 F 11.4
16:00 228.7 F 6.9
18:00 240.9 F 11.9
20:00 237.6 F 9.6
22:00 239.7 F 10.0
24:00 247.9 F 14.4
26:00 259.0 F 20.5
28:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.2
2:00 271.5 F 29.0
3:00 272.4 F 29.0
4:00 271.5 F 29.1

Min 271.2 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.9 F 0.0
5:00 249.6 F 0.0
10:00 201.5 F 0.0
15:00 191.3 F 0.0
20:00 218.8 F 0.0
25:00 227.3 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:08
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

08/25/2023 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____