

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 008218  
15:08 16/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:32  
STERILIZING 3:29  
136.0 C 220kPa 3:29  
135.9 C 221kPa 3:59  
136.4 C 221kPa 4:29  
136.4 C 223kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 224kPa 5:59  
136.4 C 224kPa 6:29  
136.4 C 223kPa 6:59  
136.4 C 223kPa 6:59  
VENTING 6:59  
AIR DRYING 7:57  
CYCLE COMPLETE 8:11  
-----

Digital Signature #  
3BF4AEDFBDD7128F  
-----