
Midmark M11 - v1.0.0
Total Cycles: 1276

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2021 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.6 F 0.0
2:00 102.6 F 0.0
4:00 114.8 F 0.0
6:00 139.0 F 0.1
8:00 175.3 F 0.2
10:00 195.7 F 3.1
12:00 207.8 F 6.1
14:00 224.2 F 9.4
16:00 233.9 F 13.2
18:00 234.1 F 8.9
20:00 243.8 F 12.1
22:00 242.6 F 11.7
24:00 240.1 F 10.3
26:00 253.1 F 16.6
28:00 264.0 F 23.0
30:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.8 F 28.8
2:00 271.8 F 28.1
3:00 272.0 F 28.5
4:00 272.6 F 28.6

Min 271.3 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 236.2 F 0.0
10:00 221.2 F 0.0
15:00 224.2 F 0.0
20:00 211.4 F 0.0
25:00 208.0 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:33
CYCLE COMPLETE

10/19/2021 10:12 AM
ACCEPTED: X
REJECTED:
Operator: U01