

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3579

Name: RITTER2  
Sterilizer ID: V2313228  
05/01/2024 04:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 170.0 F 4.1  
4:00 195.9 F 9.7  
6:00 216.8 F 5.9  
8:00 232.5 F 11.3  
10:00 233.6 F 9.4  
12:00 236.8 F 9.1  
14:00 249.9 F 15.9  
16:00 262.5 F 23.3  
18:00 271.9 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.7 F 29.8  
2:00 271.3 F 29.1  
3:00 272.5 F 30.0  
4:00 271.9 F 29.3  
  
Min 271.2 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 209.6 F 0.0  
10:00 214.2 F 0.0  
15:00 209.8 F 0.0  
20:00 206.9 F 0.0  
25:00 204.0 F 0.0  
30:00 204.8 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 18:08  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:04  
CYCLE COMPLETE

05/01/2024 05:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_