
Midmark M11 - v1.0.0
Total Cycles: 1355

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 187.4 F 3.5
4:00 197.9 F 7.6
6:00 226.2 F 12.8
8:00 228.7 F 7.7
10:00 240.0 F 12.6
12:00 237.6 F 9.8
14:00 239.3 F 8.9
16:00 248.3 F 14.3
18:00 259.8 F 20.7
20:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.3
3:00 272.3 F 28.7
4:00 271.3 F 28.2

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 209.0 F 0.0
10:00 203.4 F 0.0
15:00 204.7 F 0.0
20:00 207.3 F 0.0
25:00 209.6 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

11/10/2021 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____