
Midmark M11 - v1.0.5
Total Cycles: 878

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 158.5 F 3.0
4:00 167.9 F 6.9
6:00 193.9 F 11.8
8:00 212.7 F 6.0
10:00 228.9 F 11.2
12:00 231.3 F 8.2
14:00 240.6 F 11.3
16:00 244.6 F 13.2
18:00 257.1 F 19.9
20:00 268.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.3 F 30.1
2:00 272.2 F 29.6
3:00 271.2 F 29.1
4:00 271.4 F 29.8

Min 271.0 F 29.1
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 227.5 F 0.0
10:00 228.1 F 0.0
15:00 230.3 F 0.0
20:00 226.8 F 0.0
25:00 226.7 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

03/14/2022 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____