
Midmark M11 - v1.0.0
Total Cycles: 4647

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	158.7 F	0.0
2:00	157.6 F	2.1
4:00	171.7 F	5.5
6:00	184.6 F	8.9
8:00	202.8 F	13.1
10:00	214.8 F	5.2
12:00	225.3 F	9.2
14:00	237.0 F	13.5
16:00	237.3 F	10.2
18:00	238.1 F	8.8
20:00	247.0 F	13.8
22:00	257.8 F	19.6
24:00	267.8 F	25.9

mm:ss	Degrees	PSI
0:00	271.5 F	28.4
1:00	272.6 F	29.2
2:00	271.8 F	28.5
3:00	271.7 F	28.8
4:00	272.4 F	29.1

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	210.6 F	0.0
5:00	195.5 F	0.0
10:00	200.4 F	0.0
15:00	198.4 F	0.0
20:00	193.4 F	0.0
25:00	199.7 F	0.0
30:00	203.3 F	0.0

Drying Complete

FILLING: 0:56
HEATING: 25:27
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:08
CYCLE COMPLETE

06/17/2024 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____