

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1793

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2022 01:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
2:00 183.8 F 0.3  
4:00 210.2 F 2.2  
6:00 222.7 F 7.2  
8:00 233.0 F 12.4  
10:00 232.8 F 9.4  
12:00 235.8 F 9.5  
14:00 241.5 F 11.6  
16:00 247.2 F 14.2  
18:00 260.2 F 21.5  
20:00 271.9 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.2 F 29.3  
2:00 271.9 F 29.5  
3:00 272.0 F 29.1  
4:00 271.9 F 29.5

Min 271.4 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 218.3 F 0.0  
10:00 224.6 F 0.0  
15:00 227.1 F 0.0  
20:00 223.4 F 0.0  
25:00 226.7 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 20:31  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

11/30/2022 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_