
Midmark M11 - v1.0.0
Total Cycles: 1504

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2022 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 185.8 F 3.3
4:00 195.5 F 7.1
6:00 214.1 F 11.7
8:00 224.6 F 4.9
10:00 236.4 F 9.9
12:00 235.9 F 8.3
14:00 243.4 F 11.9
16:00 239.5 F 9.9
18:00 252.3 F 16.2
20:00 263.4 F 22.6
22:00 272.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.0 F 28.3
2:00 271.8 F 28.7
3:00 272.5 F 28.7
4:00 272.6 F 29.2

Min 271.0 F 27.8
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 235.6 F 0.0
10:00 232.3 F 0.0
15:00 226.6 F 0.0
20:00 226.9 F 0.0
25:00 230.5 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

01/10/2022 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____