
Midmark M11 - v1.0.5
Total Cycles: 480

Name: RITTER2
Sterilizer ID: V2313228
10/25/2021 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.4 F 0.0
2:00 106.2 F 0.9
4:00 124.8 F 3.1
6:00 144.0 F 6.1
8:00 168.9 F 10.0
10:00 193.9 F 7.3
12:00 213.6 F 7.1
14:00 226.9 F 11.4
16:00 228.7 F 7.7
18:00 241.0 F 13.0
20:00 240.8 F 11.5
22:00 252.7 F 17.5
24:00 263.4 F 23.8
26:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.7 F 29.8
2:00 271.6 F 29.3
3:00 272.8 F 29.9
4:00 271.8 F 29.2

Min 271.6 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 216.3 F 0.0
10:00 222.5 F 0.0
15:00 200.6 F 0.0
20:00 216.5 F 0.0
25:00 217.1 F 0.0
30:00 215.7 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

10/25/2021 02:01 PM
ACCEPTED: X
REJECTED:
Operator: U01