
Midmark M11 - v1.0.0
Total Cycles: 2027

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2022 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	135.2 F	0.0
2:00	134.3 F	0.0
4:00	152.8 F	0.0
6:00	178.1 F	0.0
8:00	197.3 F	0.0
10:00	211.1 F	0.0
12:00	212.1 F	0.0
14:00	212.3 F	0.0
16:00	212.4 F	0.0
18:00	212.4 F	0.0
20:00	212.4 F	0.0
22:00	212.4 F	0.0
24:00	212.3 F	0.0
26:00	212.3 F	0.0
28:00	212.4 F	0.0

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 217F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

06/16/2022 08:14 AM