

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 000530  
14:02 11/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:32  
STERILIZING 3:27  
136.0 C 227kPa 3:27  
135.9 C 230kPa 3:57  
136.2 C 231kPa 4:27  
136.0 C 229kPa 4:57  
136.1 C 230kPa 5:27  
136.2 C 230kPa 5:57  
136.2 C 231kPa 6:27  
136.3 C 232kPa 6:57  
136.3 C 232kPa 6:57  
VENTING 6:57  
AIR DRYING 7:44  
CYCLE COMPLETE 22:44  
-----

Digital Signature #  
357398C0D14110CE  
-----