

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1895

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2022 01:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
2:00 209.7 F 3.4  
4:00 211.6 F 8.6  
6:00 223.5 F 9.1  
8:00 233.3 F 9.8  
10:00 236.9 F 8.5  
12:00 244.7 F 13.5  
14:00 244.0 F 12.4  
16:00 257.3 F 19.5  
18:00 269.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.2  
2:00 271.5 F 28.4  
3:00 272.3 F 29.3  
4:00 272.2 F 28.8  
  
Min 271.2 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 241.9 F 0.0  
10:00 229.2 F 0.0  
15:00 229.1 F 0.0  
20:00 227.8 F 0.0  
25:00 229.0 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 3:41  
HEATING: 19:05  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

05/02/2022 02:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_