
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001888
9:15 25/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:10
STERILIZING 3:09
136.4 C 223kPa 3:09
136.1 C 223kPa 3:39
136.2 C 223kPa 4:09
136.4 C 225kPa 4:39
136.4 C 225kPa 5:09
136.4 C 223kPa 5:39
136.0 C 222kPa 6:09
136.4 C 226kPa 6:39
136.4 C 226kPa 6:39
VENTING 6:40
AIR DRYING 7:31
CYCLE COMPLETE 32:22

Digital Signature #
EF5796026995A208
