

-----  
Midmark M11 - v1.0.0  
Total Cycles: 733

Name: RITTER 1  
Sterilizer ID: V2178198  
03/18/2021 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.0 F 0.0  
2:00 137.1 F 0.6  
4:00 150.7 F 2.7  
6:00 177.6 F 6.5  
8:00 200.8 F 11.4  
10:00 215.1 F 4.7  
12:00 227.4 F 9.9  
14:00 233.0 F 8.3  
16:00 242.0 F 12.4  
18:00 240.5 F 10.7  
20:00 254.5 F 17.5  
22:00 266.1 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.9 F 29.4  
2:00 271.8 F 29.0  
3:00 272.3 F 28.8  
4:00 272.6 F 29.4

Min 270.9 F 28.1  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
5:00 210.6 F 0.0  
10:00 217.2 F 0.0  
15:00 227.1 F 0.0  
20:00 232.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:12  
HEATING: 23:36  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 23:30  
TOTAL CYCLE: 00:55:19

03/18/2021 03:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_