
Midmark M11 - v1.0.5
Total Cycles: 2565

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 184.4 F 2.1
4:00 198.5 F 6.7
6:00 214.9 F 12.5
8:00 226.6 F 8.4
10:00 237.1 F 11.1
12:00 242.1 F 12.6
14:00 242.0 F 11.9
16:00 255.6 F 18.8
18:00 267.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.9
2:00 272.4 F 29.4
3:00 271.5 F 28.9
4:00 271.4 F 29.2

Min 271.2 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 229.9 F 0.0
10:00 229.2 F 0.0
15:00 230.4 F 0.0
20:00 227.9 F 0.0
25:00 227.0 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

06/26/2023 11:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____