
Midmark M11 - v1.0.5
Total Cycles: 4078

Name: RITTER2
Sterilizer ID: V2313228
11/04/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 72.7 F 0.0
4:00 111.4 F 0.0
6:00 114.5 F 0.0
8:00 151.8 F 0.1
10:00 170.6 F 0.2
12:00 198.0 F 0.8
14:00 208.5 F 3.7
16:00 217.1 F 6.6
18:00 225.2 F 9.8
20:00 233.4 F 13.5
22:00 234.2 F 9.3
24:00 242.4 F 12.0
26:00 242.7 F 11.9
28:00 240.4 F 10.5
30:00 252.2 F 16.4
32:00 262.3 F 22.3
34:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 272.5 F 29.3
4:00 272.0 F 28.6

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 211.1 F 0.0
10:00 201.9 F 0.0
15:00 191.2 F 0.0
20:00 191.7 F 0.0
25:00 198.2 F 0.0
30:00 204.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 34:39
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:13:08
CYCLE COMPLETE

11/04/2024 11:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____