

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002223  
14:30 05/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:36  
STERILIZING 3:38  
136.4 C 225kPa 3:38  
136.1 C 223kPa 4:08  
136.1 C 223kPa 4:38  
136.4 C 226kPa 5:08  
136.3 C 225kPa 5:38  
136.3 C 225kPa 6:08  
136.2 C 223kPa 6:38  
136.3 C 224kPa 7:08  
136.3 C 224kPa 7:08  
VENTING 7:08  
AIR DRYING 8:01  
CYCLE COMPLETE 8:21  
-----

Digital Signature #  
9F58180CB9337FC5  
-----