

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4350

Name: RITTER2  
Sterilizer ID: V2313228  
02/11/2025 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 85.8 F 0.0  
2:00 86.2 F 0.4  
4:00 88.4 F 1.7  
6:00 116.7 F 3.9  
8:00 145.7 F 6.9  
10:00 172.7 F 10.1  
12:00 189.5 F 13.4  
14:00 200.5 F 4.8  
16:00 211.2 F 8.0  
18:00 222.6 F 11.6  
20:00 227.4 F 7.1  
22:00 238.0 F 11.8  
24:00 235.8 F 9.4  
26:00 247.2 F 14.8  
28:00 257.1 F 20.3  
30:00 266.7 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.8  
1:00 271.8 F 29.7  
2:00 272.2 F 29.9  
3:00 272.4 F 30.2  
4:00 272.0 F 30.2

Min 271.6 F 29.7  
Max 272.6 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.9 F 0.0  
5:00 184.2 F 0.0  
10:00 196.2 F 0.0  
15:00 196.5 F 0.0  
20:00 192.9 F 0.0  
25:00 192.9 F 0.0  
30:00 195.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 31:39  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:10:25  
CYCLE COMPLETE

02/11/2025 03:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_