
Midmark M11 - v1.0.0
Total Cycles: 3022

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 77.4 F 0.0
4:00 94.6 F 0.0
6:00 118.9 F 0.6
8:00 145.4 F 2.2
10:00 173.1 F 4.5
12:00 195.2 F 7.7
14:00 210.8 F 11.7
16:00 219.1 F 4.6
18:00 231.5 F 8.7
20:00 241.2 F 12.9
22:00 237.6 F 9.5
24:00 240.5 F 9.3
26:00 246.9 F 13.4
28:00 258.4 F 19.7
30:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.5
2:00 272.0 F 28.7
3:00 271.8 F 28.2
4:00 272.4 F 28.9

Min 271.5 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 228.4 F 0.0
10:00 219.0 F 0.0
15:00 215.6 F 0.0
20:00 214.4 F 0.0
25:00 220.1 F 0.0
30:00 232.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

03/22/2023 08:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____