
Midmark M11 - v1.0.0
Total Cycles: 3746

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 08:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 73.2 F 0.0
4:00 95.1 F 0.0
6:00 135.5 F 0.2
8:00 138.5 F 1.5
10:00 167.1 F 3.0
12:00 187.3 F 5.2
14:00 198.1 F 8.4
16:00 212.0 F 11.9
18:00 216.1 F 4.4
20:00 226.1 F 8.2
22:00 236.3 F 12.3
24:00 234.5 F 8.6
26:00 244.6 F 12.4
28:00 243.9 F 12.0
30:00 255.3 F 17.8
32:00 265.4 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.5 F 28.5
2:00 271.5 F 28.0
3:00 272.6 F 28.5
4:00 271.4 F 27.9

Min 271.4 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 200.6 F 0.0
10:00 204.4 F 0.0
15:00 208.4 F 0.0
20:00 208.9 F 0.0
25:00 213.5 F 0.0
30:00 214.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:55
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:12:05
CYCLE COMPLETE

09/11/2023 09:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____