

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001711  
14:44 07/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:38  
STERILIZING 3:43  
136.4 C 223kPa 3:43  
136.3 C 222kPa 4:13  
136.3 C 224kPa 4:43  
136.3 C 224kPa 5:13  
136.1 C 223kPa 5:43  
136.4 C 225kPa 6:13  
136.1 C 223kPa 6:43  
136.4 C 224kPa 7:13  
136.4 C 224kPa 7:13  
VENTING 7:13  
AIR DRYING 8:03  
CYCLE COMPLETE 8:05

-----  
Digital Signature #  
30E4201C0C67EC1F  
-----