
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007469
7:23 03/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:39
136.4 C 222kPa 3:39
136.1 C 221kPa 4:09
136.4 C 223kPa 4:39
136.4 C 222kPa 5:09
136.2 C 217kPa 5:39
136.2 C 222kPa 6:09
136.2 C 221kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 7:57
CYCLE COMPLETE 16:22

Digital Signature #
C3D297236B2F758E
